



Ref. Certif. No.

DE 3 - 32752

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Laboratory Equipment (HV Power Supply)
Name and address of the applicant	iseg Spezialelektronik GmbH Bautzner Landstr. 23 01454 Radeberg GERMANY
Name and address of the manufacturer	iseg Spezialelektronik GmbH Bautzner Landstr. 23, 01454 Radeberg, GERMANY
Name and address of the factory	iseg Leistungselektronik Sachsen GmbH Bautzner Landstr. 23, 01454 Radeberg, GERMANY
Ratings and principal characteristics	Input: Voltage : 100 Vac – 240 Vac Current : 1.5 A Frequency : 50/60 Hz Protection Class: Class I Supply connection: detachable cord Output: Voltage : 6 kV Current : 4 mA Power : 12 W
Customer's Testing Facility (CTF) Stage used	CTF STAGE 1
Model/type Ref.	SHR 20 20r, SHR 20 60r, SHR 40 20r, SHR 40 60r SHR 22 20r, SHR 22 60r, SHR 42 20r, SHR 42 60r
Additional information (if necessary)	Certificate DE 3 - 31500 issued 2021-04-19 is replaced by this version due to technical changes. Evaluated to National and Group Differences for EU, CA, US
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
as shown in the Test Report Ref. No. which forms part of this certificate	713202968

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 109910 0001 Rev. 01

Date, 2026-01-27

Siemon

(Thorsten Siemon)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service