




America

CERTIFICATE

No. U8 109910 0002 Rev. 00

Holder of Certificate: **iseg Spezialelektronik GmbH**
Bautzner Landstr. 23
01454 Radeberg
GERMANY

Certification Mark:



Product: **Laboratory Equipment**
(HV Power Supply)

Model(s): **SHR 20 20r, SHR 20 60r, SHR 40 20r, SHR 40 60r**
SHR 22 20r, SHR 22 60r, SHR 42 20r, SHR 42 60r

Parameters: Voltage : 100Vac – 240Vac
Current : 1.5A
Frequency : 50 / 60Hz

Tested according to: **UL 61010-1:2012/R:2018-11**
CAN/CSA-C22.2 No. 61010-1:2012/A1:2018-11

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 713202968

Date, 2021-04-16

(Abdul Sabbagh)